VEHICLE CALIBRATION CERTIFICATE

Client: Ampthill Metal

Model: RC2069

Detector SN: **22184 & 22185**

Control Panel SN: 1189REC4

System Controller: RadLink

TeamViewer ID: 388 922 445

Software: **5.0 (7116)**

Date Tested: 18/03/2024 Valid Until: 31/03/2025

BG Count: Detector 1: 3077 Detector 2: 3737

Test 1:

Result **OK**

LED Test: Detector 1: 23415 Detector 2: 24692

Test 2:

Result OK

Source Test: Detector 1: 4536 Detector 2: 5887

Source: *0.25uCi Cs137*

Comments:

Serviced By: Tara Higgins

Signature:

This document certifies that the above equipment has been tested and verified to be functioning as per RadComm Systems specifications.





